

Model P813 PH /Conductivity Meter

Operation Manual

Table of Contents

1. Brief Introduction -----	3
2. Technical Parameters -----	4
3. Instructions to the Meter -----	6
3.1. Main Interface Display -----	6
3.2. PH Interface Display -----	10
3.3 Conductivity Interface Display-----	15
3.4. The Storage, Recall and Elimination of the Measuring Information -----	19
4. P/H Measurement -----	19
4.1. Preparation Work -----	19
4.2. Meter Calibration -----	19
4.3. PH solution Test -----	21
4.4. Considerations-----	21
5. MV Measurement -----	23
6. Conductivity Measurement -----	23
6.1. Preparation Work -----	23
6.2. Meter Calibration -----	23
6.3. Conductivity solution Test -----	23
6.4. Important Statement -----	24
6.5. Parameter Setting -----	25
6.6. Considerations -----	25
7. Meter's Complete Kit -----	26
8. Warranty -----	26
Appendix A conductivity standard solution concentration and conductivity value	

1. Brief Instruction:

Thanks for buying and using the model P813 pH / mV / Conductivity Meter (the following called “meter” in short).

Before using this meter, please read the operation manual carefully in order to help use and maintain it correctly. On the basis of improving instrument of performance constantly, we reserve the right of changing the content of this manual and accessories in case of not notifying in advance.

This meter is a perfect combination with the most advanced electronic technology, sensor technology and software design. The meter can measure the parameters of pH/ conductivity and temperature for high accuracy water solution. It is the best portable pH/conductivity meter with the highest performance and the lowest cost. It is suitable for the trade such as the mining industry, power plant, water treatment projects and environmental protection, etc.

Built-in ARM32 bit microprocessor chip, beautiful appearance, multi-functional and easy to use, this meter has the following prominent features:

- 1.1. Built-in microprocessor chip, with the intelligent functions of automatic calibration, automatic/manual temperature compensation, data storage, multiple function settings.
- 1.2. Adopts digital filter and step slipping technology to intelligently improve meter's response speed and result accuracy.
- 1.3. The new pH meter electrode and the temperature electrode are used to make the PH measurement model have automatic temperature compensation and manual ammonia compensation function. It is more convenient to use, more accurate and more accurate, new electrical conductivity electrode and temperature electrode are used at the same time. The electrical conductivity measurement model has automatic temperature

compensation function, and it can also measure RES, TDS, and also can be used. SAL is more convenient to use and more accurate.

1.4. PH calibration can automatically identify 15 kinds of buffer solutions. There are three standard buffer solutions: the European and American series, the NIST series and the China series, can be calibrated at one point, two point and three point, and automatically identify 8 conductivity calibration solutions when the electrical conductivity is calibrated. There are two kinds of standard liquid: the European and American series and the Chinese series. The point calibration method is convenient for the customer to choose.

1.5. Meter's circuit board adopts SMT film-covering technology to improve meter's production reliability.

1.6. Adopt 1024*600 high resolution 7 inch TFT color LCD screen, capacitive touch screen, and operate more sensitively.

1.7 The meter has wireless Bluetooth function and can connect to a wireless Bluetooth printer or connect to mobile phone APP..

1.8. Dustproof and waterproof meter meets IP57 rating.

2. Technical Parameters:

2.1. pH:

Measuring range	(-2.00 - 19.999) pH
Resolution	0.1/0.01 pH
Accuracy	Meter: ± 0.002 pH; Complete Kit: ± 0.01 pH
Input current	$\leq 1 \times 10^{-12}$ A
Input impedance	$\geq 3 \times 10^{12} \Omega$
Stability	± 0.01 pH/3h
Temp. Compensation range	(0 - 100) °C (auto/manual)

2.2. mV:

Measuring range (mV/ORP/E _H)	-1999.9 mV - 0 - 1999.9mV
Resolution	0.1mV
Accuracy	Meter:±0.3% FS

2.3. Conductivity:

Measuring range	Conductivity: (0.00 - 19.99) μ S/cm (20.0 - 199.9) μ S/cm (200 - 1999) μ S/cm (2.00 - 19.99) mS/cm (20.0 - 199.9) mS/cm TDS: (0 - 100) g/L Salinity : (0 - 100) ppt Resistivity: (0 - 100) M $\Omega \cdot$ cm
Resolution	0.01/0.1/1 μ S/cm 0.01/0.1 mS/cm
Accuracy	Meter:±0.5% FS, connect meter with electrode:±1.0% FS
Temp. compensation range	(0 - 100) $^{\circ}$ C (manual / auto)
Electrode constant	0.1 / 1 / 10 cm ⁻¹
Benchmark temperature	25 $^{\circ}$ C, 20 $^{\circ}$ C and 18 $^{\circ}$ C

2.4. Other Technical Parameters:


Data storage	300 groups
Storage content	date, measuring value, measuring unit and temperature,
Power	DC12V/1A
Size and weight	Meter: (165×90×32)mm/310g
Quality and safety certification	ISO9001:2000, CE and CMC

2.5. Working Condition:


Environment temperature	5 - 35 $^{\circ}$ C (0.01 grade)
Environmental humidity	≤85%
IP rating	IP54 Dustproof and waterproof

3. Instructions to the Meter:


3.1. Main Interface Display:




3.1.1 Setting (by setting the key, as in the following diagram)




(a) Time setting: setting the date and time menu, modifying the current clock time; setting the display according to the time, such as the following figure, changing the date after the arrow at the back of the date and time, by determining the date and time to save the modified date, by cancellate.




(b) Language settings: Chinese and English settings; set the display in language as below, select CH or EN according to the required display interface, by determining the save settings, by cancellations.




(c) Brightness adjustment: adjust the LCD backlight brightness according to the brightness adjustment, as shown below, drag the delimit to set the brightness, press the confirmation save settings to cancel the return.



(d) Temperature unit: centigrade or optional (defaults); according to the temperature unit displayed as the following figure, select the unit to compensate for the temperature of the unit C or to determine the storage setting, according to the cancellation return.




(e) Bluetooth settings: connect to Bluetooth printer or mobile phone APP. According to the Bluetooth settings, you can choose to connect the Bluetooth printer or the phone APP, enter the authorization code, press the confirm save settings, press cancel to return.



(f) Exit settings: return to main interface after pressing exit settings.

3.1.2 Equipment about

Display information about device hardware and software version as shown below.



3.1.3 Printing


Print the current measurement parameters according to the measurement.

3.1.4 Sleep

Turn off LCD backlight display according to sleep device and enter sleep standby status as shown below.




3.2. PH interface display: (at the main interface of the instrument displayed by PH as shown below)



3.2.1. Display interface description

- (1).PH measurement data display interface;
- (2).MV measurement data display interface;
- (3).Measuring temperature shows that "A" stands for automatic, and "M" stands for manual operation.
- (4).The calibration indicator of the standard buffer solution;
- (5).The indicator of the state of the electrode.
- (6). Storage equipment sign
- (7).Sign of stability

3.2.2. Settings: enter the settings menu as set by setting




- (a) Adding ammonia compensation: "no compensation", "pure water compensation", "ammonia compensation" three modes; according to the addition of ammonia compensation into the following diagram, according to the determination of the storage settings, by cancellations return.

PH						2018/4/10 9:54
SET			Nh3		No cal	
Nh3	Clear	Tmp Src	NONE			
Std	Res	Tmp	PURE H2O			
Reset	Export	Quit	NH3 H2O		★★★★★	
			Cancel		Ok	
SET	QUERY	CAL	SAVE	PRINT	BACK	


(b) Record empty: empty the current storage of all data. Clear the entries according to the record, clear the empty storage records, and cancel the return.

PH						2018/4/10 9 54
SET			Clear		No cal	
Nh3	Clear	Tmp Src	Record num			
Std	Res	Tmp	11			
Reset	Export	Quit	Cancel		Clear	
SET	QUERY	CAL	SAVE	PRINT	BACK	

(c) The source of temperature: "automatic" and "manual" two modes; according to the temperature source, such as the following diagram, the selection is or automatically press the set to save, by cancellation return.



(d) Series of standard liquid: standard calibration liquid, "national standard", "NIST" and "European and American" series; according to the series of standard liquid, such as the following diagram, the selection standard is determined to save the setting and is returned by cancellation.



(e) Resolution: 0.1pH and 0.01pH; according to the resolution, select the appropriate resolution as shown below, press the confirmation save settings and cancel the return.

PH						2018/4/10 9:55
SET			Res			No cal
Nh3	Clear	Tmp Src	0.1			 ★★★★★
Std	Res	Tmp	0.01			
Reset	Export	Quit	Cancel Ok			
SET		QUERY	CAL	SAVE	PRINT	BACK

- (f) Manual temperature: manual temperature setting;
- (g) Resume factory: restore factory settings, delete all records;
- (h) Data export: save the data into the U disk.
- (i) Quit Setup: return to the test interface by pressing the exit settings.

3.2.3. Query

Querying all current saved data, including information, mode, temperature, and date of preservation. According to the query, display the following page, press the previous page, turn the next page, delete the data by clear, and exit according to the exit.

PH						2018/4/10 9:54
Record query						No cal
NO	PH	MV	TMP	DATE		 ★★★★★
1	7.00 pH	0 mV	25.0 °C	2018/3/29 14:45		
2	7.00 pH	0 mV	25.0 °C	2018/3/29 14:47		
3	7.00 pH	0 mV	25.0 °C	2018/3/29 14:49		
4	7.00 pH	0 mV	25.0 °C	2018/3/29 14:51		
5	7.00 pH	0 mV	36.7 °C	2018/3/30 10:50		
6	7.000 pH	0 mV	25.0 °C	2018/3/30 13:6		
7	7.00 pH	0 mV	25.0 °C	2018/3/30 16:5		
8	7.00 pH	0 mV	25.0 °C	2018/3/30 16:5		
Clear		Pre page	1/2	Next page	Quit	
SET	QUERY	CAL	SAVE	PRINT	BACK	

3.2.4. Calibration

Using standard calibration liquid to calibrate the electrode, the instrument can be calibrated at one point, two points and three point three kinds. When the user is calibrated, it can be calibrated at any point. No order is needed. After the calibration, the standard value of the calibrated solution will be displayed on the right side to facilitate the customer's view.


3.2.5. Preservation

Save the current measurement value, including information on measurement, mode, temperature, and date of preservation.

3.2.6. Printing

To print the current measurement data, the instrument can connect the wireless Bluetooth printer. Before printing, you need to set the Bluetooth as the printer mode (the default is the printer mode) in the system settings, open the printer power, the printer is close to the device, and directly click the "print" button to print the current measurement value.


3.3. COND interface display: (at the main interface of the instrument displayed by COND, as shown below)



3.3.1. Display interface description

- (1) Measurement data display interface;
- (2) Mode selection, "COND", "RES", "TDS" and "SAL" are four modes.
- (3) Measuring temperature shows that "A" stands for automatic, and "M" stands for manual operation.
- (4) Setting of the electrode constant (default 1),
- (5) TDS coefficient setting (default 0.50).
- (6) Storage equipment sign
- (7) Sign of stability

3.3.2. Settings: enter the settings menu as set by setting



- (a) Reference temperature: "18 degrees", "20 degrees", "25 degrees", three modes; at the base temperature, enter the following diagram, select the appropriate reference temperature to determine the storage settings, by cancellations return

COND						2018/4/10 9 55
SET			Ref tmp		Pole const	
Ref tmp	Clear	Tmp src	18 °C		1.000	
Std	Tmp coe	Tmp	20 °C		TDS coe	
			25 °C		0.50	
Reset	Export	Quit	Cancel		Ok	
SET	QUERY	CAL	SAVE	PRINT	BACK	

(b) Record empty: empty all stored data. Clear the entries according to the record, clear the empty storage records, and cancel the return.

COND						2018/4/10 9:56
SET			Clear		Pole const	
Ref tmp	Clear	Tmp src	Record num		1.000	
Std	Tmp coe	Tmp	3		TDS coe	
					0.50	
Reset	Export	Quit	Cancel		Clear	
SET	QUERY	CAL	SAVE	PRINT	BACK	

(c) The source of temperature: "automatic" and "manual" two modes; according to the temperature source, such as the following diagram, the selection is or automatically press the set to save, by cancellation return.

COND						2018/4/10 9 56
SET			Tmp Src		Pole const	
Ref tmp	Clear	Tmp src	☒ AUTO ☐ MANUAL		1.000	
Std	Tmp coe	Tmp			TDS coe	
Reset	Export	Quit	Cancel		Ok	
SET	QUERY	CAL	SAVE	PRINT	BACK	

(e) Series of standard liquid: standard calibration liquid, "national standard", and "European and American" series; according to the series of standard liquid, such as the following chart, the selection standard is determined to save the setting and is returned by cancellation.

COND						2018/4/10 9 56
SET			Std		Pole const	
Ref tmp	Clear	Tmp src	☒ CH ☐ USA		1.000	
Std	Tmp coe	Tmp			TDS coe	
Reset	Export	Quit	Cancel		Ok	
SET	QUERY	CAL	SAVE	PRINT	BACK	

- (e) Temperature compensation coefficient: input the value to be compensated according to the temperature compensation coefficient (0.00-9.99%).
- (f) manual temperature: manual temperature setting at 0.0-99.9 C;
- (g) resume factory: restore factory settings, delete all records;

(h) data export: save the data into the U disk.

(i) quit Setup: return to the test interface by pressing the exit settings.

3.3.3. Inquire

Querying all current saved data, including information, mode, temperature, and date of preservation. According to the query, display the following page, press the previous page, turn the next page, delete the data by clear, and exit according to the exit.

COND

2018/4/10
9:56

Record query

NO	VALUE	MODE	TMP	DATE
1	0.000 μ S/cm	COND	25.0 $^{\circ}$ C	2018/3/29 14:45
2	0.000 μ S/cm	COND	25.0 $^{\circ}$ C	2018/3/30 11:54
3	0.000 μ S/cm	COND	25.0 $^{\circ}$ C	2018/3/30 16:30

Pole const
1.000

TDS coe
0.50

Clear

Pre page

1/1

Next page

Quit



SET



QUERY



CAL



SAVE



PRINT



BACK

3.3.4. calibration

The standard calibration liquid is used to calibrate the electrode, the electrode is put into the standard solution, the stable symbol is set from red to green, and the new electrode has been calibrated when it is cancelled and the new electrode is in the factory. Its constant value is marked on the electrode, and the calibration is not needed again, and the direct input electrode is often used. The value can be used as the following diagram.

COND					
Please put std solution ● Cancel Ok					2018/4/10 9 56
CAL					Pole const 1.000
← MODE COND →			A 25.0°C	TDS coe 0.50	
SET	QUERY	CAL	SAVE	PRINT	BACK

3.3.5. preservation

Save and save the current measurement value, including the measurement value, mode, temperature value, and date of preservation.

3.3.6. printing

By printing and printing the current measurement data, the instrument can connect the wireless Bluetooth printer. Before printing, it is necessary to set the Bluetooth as the printer mode in the system setting (the default is the printer mode), open the printer power, the distance from the printer to the device as close as possible, and directly click the "print" button to print the current. The measured value.

3.4. Measurement information preservation, query and delete:

3.4.1. Save measurement information:

In the measurement mode, when the measured value is stable, the upper right corner stability symbol is changed from red to red.

For green, press the Save button to save the current measurement value. The instrument

can store 300 sets of numbers.

According to it.

3.4.2. Query measurement information:

In the measurement mode, click the "query" button to query the saved data, and the data information contains the measurement value, the mode, the temperature value, and the storage date.

Support page flip function, each page contains 8 saved data, press "last page" or "down".

Page "page and check.

3.4.3. Removes the stored measurement information:

In data query mode, click the clear button to confirm whether the system is cleared.

All records can be deleted again by clicking "confirm".

4. pH Measurement:

4.1. Preparation Work:

4.1.1. Switch on the DC12V/1A power and press the "pH" button at the main interface to enter the pH measurement interface.


4.1.2 Check E-201-9 composite eletrode, check whether the PH composite electrode glass remains moist. If the glass bubble is damaged, the electrode will not be used properly. The surface of the bulb is too dry. It needs to be soaked in saturated potassium chloride solution for 24 hours, then activated, and then used.

4.1.3.Insert the PH composite electrode and the temperature electrode into the corresponding interface.

4.2. Meter Calibration:

4.2.1. Selection of standard calibration liquid.

- (a) Select the "standard solution series" in the settings menu. The three series are selected as follows.




Three series of standard buffer solution values:

CH (China Series): 1.68 pH, 4 pH, 6.86 pH, 9.18 pH, 12.46pH

NIS (NIST Series): 1.68 pH, 4.01 pH, 6.86 pH, 9.18 pH, 12.45pH

USA (European and American Series): 1.68 pH, 4 pH, 7 pH, 10.01 pH, 12.45pH

- (b) After the standard is selected, enter the calibration mode according to the "calibration" button (as shown below).



The specific calibration steps are shown in the following calibration instructions.

4.2.2. Calibration description

- (a) The instrument can be automatically calibrated at one point, two or three points. When the first calibration is completed, the calibration mode is pulled out and the calibration mode is entered. The right display area will show the calibrated standard liquid value. When the measurement accuracy is less than $+0.1\text{pH}$, a buffer solution is selected according to the range of measurement. The point calibration is OK.
- (b) When the second point calibration is completed, press "Cancel" to exit the calibration mode and enter. In the measurement mode, the right display area shows the calibrated standard liquid value, the measured value is only in the acid range, and the $\text{pH}4.00$ and $\text{pH}6.86$ calibration can be selected; if the measurement is only in the alkaline range, the calibration of $\text{pH}6.86$ and $\text{pH}9.18$ can be selected.
- (c) If the measuring range is wide, or the pH electrode has long aging time, the three point calibration should be selected, which will make the measurement accuracy higher. However, the pH electrode used for the first time must be calibrated at three points, and the slope of the adjustment instrument is consistent with that of the pH electrode.

4.3. PH Solution Test:

Immerse pH electrode into the sample solution after washing and dry it, rock the electrode holder and still, When the measured value is stable and displays the green stable icon, the reading value is the pH value of the solution.

Note: According to the pH equal temperature measuring theory: the closer the temperature of the sample solution with the calibration solution, the more accurate the measuring value will be acquired. So please comply with this theory.

4.4. Considerations:

- 4.4.1. Calibration times of meter rely on the sample, electrode performance and required accuracy. For high accurate measurement ($\leq \pm 0.02\text{pH}$), which should be calibrated immediately with high accurate standard buffer solution, for general accuracy measuring ($\leq \pm 0.1\text{pH}$), which can be used almost one week or long time once be calibrated.
- 4.4.2. The meter must be recalibrated in the following situations:
- (a) New changed or unused electrode for a long time;
 - (b) After measuring acid ($\text{pH} < 2$) or alkaline ($\text{pH} > 12$) solution;
 - (c) After measuring solution which contains fluoride and concentrated organic solution;
 - (d) The solution's temperature is much different with calibration temperature.
- 4.4.3. The soaking solution in the protecting bottle of front pH electrode is to keep the glass bulb and junction activating. Loose the capsule, pull out the electrode and wash it in purified water before measuring. Insert the electrode and screw tight the capsule after measuring to prevent the solution leaking. If the soaking solution is turbid or moldy, please wash and change a new one at once.
- 4.4.4. The preparation of the soaking solution: take 25g analytical pure KCL, dissolved with purified water and dilute to 100mL. Electrode should avoid soaking in purified water protein solution and acid fluoride solution for a long time as well avoid getting

touching with organic silicon lipidic matters.

- 4.4.5. To calibrate the meter with the given value pH buffer solution, the pH value of the standard buffer solution must be reliable so as to improve the accuracy. Buffer solution should be changed in time after many times using.
- 4.4.6. Always keep the meter clean and dry; especially for the socket of meter and electrode, otherwise it may lead to an inaccurate measurement or invalidity. To clean and dry them with medical cotton with dehydrated alcohol if there are any dirty.
- 4.4.7. The sensitive glass bulb in the front of combination electrode should not touch with hard things, any broken and rough will make the electrode invalidity. Before and after measuring, the electrode should be washed with purified water, and dry electrode after washing, don't clean glass bulb with tissue for it will effect stability of electrode potential and enlarge response time. The electrode should be washed many times for removing the sample stuck on the electrode, or wash with suitable solvent then clean the solvent with purified water after measuring sticky sample.
- 4.4.8. An electrode be used for a long time, or measured solution which contains a polluting solute easily for the sensitive bulb, or a substance resulting in jam at the junction, the electrode will be getting passivated, its sensitivity will decrease and its response is getting slow, the reading are not correct. It could adopt the following method for various cases:
- (a) The glass bulb is contaminated and aging: Put the electrode into 0.1mol/L dilute hydrochloric acid (Preparation: diluted 9mL hydrochloric acid to 1000mL with purified water) for 24h. Rinse it with purified water, then dipped it into the electrode dipping solution for 24h. If the passivation is serious, then user can also put the bulb of electrode into 4% HF (hydrofluoric acid) or the electrode activation solution for 3 to 5 seconds, rinsing it with purified water, and dipped it in the electrode soaking solution for 24h to renew it.

(b) Wash for contaminated glass bulb and junction: (For reference)

Contamination	Abluent
Inorganic metal oxide	diluted acid less than 1mol/L
Organic lipidic matter	dilute washing (weak alkaline)
Resin macromolecule matter	dilute alcohol, acetone, ether
Proteinic haematocyte sediment	Acidic enzymatic solution (such as dried yeast)
Kinds of paint	dilute bleach, peroxide

4.4.9. pH electrode using period is about 1 year, but its life will be shortened if using condition is poor or incorrect maintenance. So it should be replaced immediately after electrode become aging or invalid.

4.4.10. When the meter is abnormal, please restore the device to the factory settings, then calibrate and test it.

5. mV Measurement:

5.1. Switch on the DC12V/1A power and press the "pH" button at the main interface to enter the pH measurement interface.

5.2. Is connected to the ORP composite electrode or ion composite electrode. Wash and dry, immerse in the tested solution, stir it slightly, then place it at rest. When the green stability indicator icon is stabilized and displayed, the current measurement is displayed in the mV display area.

5.3. When mV is measured, the customer only needs to read the values in the mV display area, and the display values in the pH display area. It can be ignored and only read the mV display value.

6. Conductivity Measurement:

6.1. Preparation Work:

6.1.1. Switch on the DC12V power supply and press the "COND" button to select the conductivity measurement mode.

6.1.2. Check whether the DJS-1.0 conductivity electrode pole piece is intact. If the electrode pole piece glass is damaged, the conductivity electrode pole piece rusts will cause the electrode not to be used normally, so we should replace the new electrode.

6.1.3. Connect the DJS-1 conductivity electrode and temperature electrode into the meter.

6.2. Meter's Calibration (Standard solution calibration method):

According to the "calibration" key, enter the calibration mode, clean the DJS-1 conductance electrode with pure water and dump it dry, immerse in the 1408 S/cm calibration solution with the temperature electrode, stir it at rest, wait for stability, click "confirm", and then show the standard value of the current calibration solution at this time, 1408 μ S/cm, and then confirm the confirmation again. Calibration results, calibration is completed, exit calibration mode, enter the measurement mode.

After calibration, the normal value of the current calibration electrode will be displayed on the right side of the display area.

Note: (a) The meter has been calibrated before leaving factory, so generally users can use it directly.

(b) The standard solution method is aimed at the long time use of conductance electrode, which leads to inaccuracy of conductance constant. The new conductance electrode has been calibrated before being released. Its constant is marked on the electrode and input constant before use.

6.3. Conductivity solution Test:

6.3.1. Clean and dry the conductance electrode, put it in a solution with the temperature electrode, stir it at rest, wait for the data to be stable, and read the data, that is, the measurement of the conductivity of the solution.

6.3.2. Clicks "MODE" or "to" in the "MODE" area, which can display conductivity value, resistivity value, salinity value and salinity value.

6.4. Important Statement:

6.4.1. This meter build-in below two kinds of calibration solution series,

(a) **USA** (Europe & U.S.A. series) — 84 μ S/cm, 1413 μ S/cm, 12.88 mS/cm and 111.9 mS/cm

(b) **CH** (China series) — 146.6 μ S/cm, 1408 μ S/cm, 12.85mS/cm and 111.3 mS/cm

6.4.2. This meter has an unique one-point calibration function, to choose one calibration solution based on the principle of the water samples and calibration solution conductivity as close as possible, in general the most common calibration solution is 1408 μ S/cm. Use the equipped DJS-1 conductivity electrode (K=1), and do calibrate with 1408 μ S/cm calibration solution, can use within the range of less than 100 mS/cm. Please choose refer to the chart (6-1).

Chart (6-1)

Measuring range	0.05 to 20 μ S/cm	0.5 μ S/cm to 200mS/cm		
Electrode constant	K=0.1cm ⁻¹ (flow test)	K=1.0cm ⁻¹		
Calibration solution	84 μ S/cm	84 μ S/cm	1413 μ S/cm	12.88 mS/cm 111.9 mS/cm

6.4.3. There are two kinds of method which are standard solution calibration method and constant calibration method for conductivity electrode calibration which have been set in meter, the statement in item 4.2. “electrode calibration” is the standard calibration method, and when the accuracy of standard solution is accurate then suggest the standard solution calibration method is priority selected, it can ensure the best accuracy. If user used to use constant calibration method, it can be set according to the

constant on the conductance electrode (The new conductance electrode manufacturers have been standard accurate, customers can be used at ease, such as long-term no use or fouling. In order to ensure accuracy, first clean up the electrode and then calibrate the standard solution.)

6.4.4. The temperature compensation coefficient of the meter setting is 2.00%/°C by producer. However, the conductivity temperature coefficient is different for solution of different variety and concentrations, the user can refer chart (6-2), as well as the data which they own get in the experiment, to set in the parameters setting of parameter setting. At the same time, meter will do automatic non-linear temperature compensation in the high purified water which less than 10 μ S / cm.

Note: When the coefficient of temperature compensation is set to be 0.00, that is, there is no temperature compensation when testing, the measuring value based on current temperature.

Chart (6-2)

Solution	Temperature compensation coefficient
NaCl salt solution	2.12%/°C
5%NaOH solution	1.72%/°C
Dilute ammonia solution	1.88%/°C
10% hydrochloric acid solution	1.32%/°C
5% sulfuric acid solution	0.96%/°C

6.5. Considerations:

- 6.5.1. The conductivity electrode has been calibrated before meter leaving factory, its constant value is marked on the electrode, user can direct set constant value, direct use without calibration.
- 6.5.2. Recommend calibrating one time every month under the normal circumstances; The use of the conductance electrode after a period of time also needs to be calibrated once.
- 6.5.3. Keeping the conductivity electrode clean and wash it with purified water, then throw off the water on it before and after testing. It's better to rinse electrodes with sample solution;
- 6.5.4. The DJS series platinum black conductivity electrode surface is plated with a layer of metal platinum black to be used as lowering the electrode polarization, enlarge measuring range, so can not polish the surface of platinum black, only can wash it in water by shaking, in case to damage the platinum black coating; for organic stained things can be washed with hot water with detergent or alcohol.
- 6.5.5. Conductivity electrode before use can be immersed in purified water, to prevent the platinum black from passivation, if found the platinum black plated electrode is invalid, can immerse it into 10% nitric acid solution or 10% hydrochloric acid for 2 minutes, then rinse with purified water and again test. If the situation does not improve, then need to replace platinum black, or replacing a new conductivity electrode.
- 6.5.6. When the meter is abnormal, please set up the instrument to restore the factory settings, then calibrate and test it.

7. Meter's Complete Kit:

7.1. Model P813 pH/ Conductivity meter	1unit
7.2. E-201-9 composite electrode	1pc
7.3 DJS-1 Conductivity electrode	1pc
7.4. T-10-Q temperature electrode	1pc
7.5. Standard buffer solution (4.00, 6.86, 9.18)	2set
7.6.12V/1A power adapter	2 pc
7.7. Manual	1pc
7.8 .Electrode bracket(containing base)	1pc

8. Warranty:

- 8.1. We warrant this meter to be free of charge maintain, replace the parts or products under normal using circumstances, from purchased time within one year caused by manufacturing bad and unable to work.
- 8.2. Attached electrodes do not belong to this warrant range. But, if the newly purchased electrode went wrong without using, it's free of charge to maintain and replace.
- 8.3. The above warranty is not apply to defects resulting from action of user such as misuse, improper wiring, operation outside of specification, improper maintenance or repair, or unauthorized modification.

Appendix A

Standard solution concentration and electrical conductivity of electrical conductivity

Solution number	Reference solution KCLg/1000g solution (In vacuum)	Reference solution KCLg/1000g solution (20 °C room temperature)	Conductivity/(S.cm ⁻¹)				
			15°C	18°C	20°C	25°C	35°C
1	71.1352	74.2457	0.09212	0.09780	0.10170	0.11131	0.13110
2	7.41913	7.4365	0.010455	0.011163	0.011644	0.012852	0.015353
3	0.745263	0.7440	0.0011414	0.0012200	0.0012737	0.0014083	0.0016876
4	0.074528	Dilute the number 3 100ml to 1000ml	0.0001185	0.0001267	0.0001322	0.0001465	0.0001765

Notes: The following conditions must be observed in the application of the above standard solution

- 1, The standard values listed in the table deduct the water conductivity of the standard solution.
- 2, The standard solution of the solid standard substance of electrical conductivity can be prepared by baking 4h at 110°C.
- 3, Configuration of standard substances according to the environmental conditions specified in Table 2.
- 4, It is recommended to use a first class 1L capacity bottle and a balance of 0.1mg.

